



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **CIRION BioPharma Recherche Inc.**, using **Bio-Rad Variant II** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: <b>Variant II</b>	Calibrator Lot: <b>64337554, 64337556</b>	Column Lot: <b>70415</b>
Reagent Lot: <b>64284285, 64284286, 64252822</b>	Calibrator Assigned Values: <b>5.4%, 10.2%</b>	

Date of Certification: January 1, 2021

Certification Expires: January 1, 2022

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor